Uniphos Detector Tube Data SheetDirect Reading Length of Stain Type

Benzene

STEL (TLV): 2.5 ppm TWA (TLV): 0.5 ppm Flammable Range: 1-8%

Specification:

Measurement Range	Extended	Standard	Extended	
	2.5 - 50	5 - 100 ppm	10 - 200	
No. of Pump Strokes	2 (200 mL)	1 (100 mL)	1/2 (50 mL)	
Volume Correction Factor (VCF)*	0.5	1.0	2	
Sampling Time	2 minutes per pump stroke (100 mL)			
Color Change	White → Brownish green			
Shelf Life	2 years			
Storage condition	Below 10°C (50°F)			
Calibration condition	Calibrated at 20°C (68°F) and 50% RH.			
Reaction Principle	$C_6H_6 + I_2O_5 + H_2SO_7 \rightarrow I_2$			
Relative standard deviation	± 10 %			

^{*} Multiply the observed reading by the correction factor (VCF) to obtain the true concentration.

Possible Interferences:

Compound	Concentration (ppm)	Interference	Colour change / Comments
Aliphatic hydrocarbon		No	
H ₂ S	50		Brown ring
Isobutylene	106	+	
Toluene		+	White to brownish green

Correction for Environmental Parameters:

Temperature	Not necessary between 0 - 40°C (32 - 104°F).	
Relative humidity	Not necessary between 10 - 90 %.	

Manufactured By: Uniphos Envirotronic Pvt. Ltd. Nahuli (PO), Umbergaon (Tal.) Valsad (Dist), Gujarat, India - 396108