Uniphos Detector Tube Data SheetDirect Reading Length of Stain Type

Trichloroethylene CI₂C: CHCI TWA (TLV): 50 ppm STEL (TLV): 100 ppm

Flammable Range: NA

Specification:

Management Dance	Extended	Standard	Extended	
Measurement Range	Do not extend	20 – 250 ppm	Do not extend	
No. of Pump Strokes		2 (200 mL)		
Volume Correction Factor (VCF)*		1.0		
Sampling Time	1.5 minutes per pump stroke (100 mL)			
Color Change	White → Yellow			
Shelf Life	1 year			
Storage condition	Below 10°C (50°F)			
Calibration condition	Calibrated at 20°C (68°F) and 50% RH.			
Reaction Principle	$Cl_2C:CHCl + Cr^{6+} + H_2SO_4 \rightarrow Cl_2$			
	Cl ₂ + o -Tolidine → Reaction product			
Relative standard deviation	± 10 %			

^{*} Multiply the observed reading by the correction factor (VCF) to obtain the true concentration.

Possible Interferences:

Compound	Concentration (ppm)	Interference	Colour change / Comments
1,2 Dichloroethylene		+	White to Yellow
Cl ₂		+	White to Yellow

Correction for Environmental Parameters:

Temperature	Not necessary between 0 - 40°C (32 - 104°F).	
Relative humidity	Not necessary between 10 - 90 %.	

SCE-4L