# **Uniphos Detector Tube Data Sheet**Direct Reading Length of Stain Type

# Chloroform

CHCI<sub>3</sub> STEL (TLV): NA

TWA (TLV): 10 ppm Flammable Range: N.A

## **Specification:**

Measurement Range	Extended	Standard	Extended
		10 – 100 ppm	
No. of Pump Strokes		5 (500 mL)	
Volume Correction Factor (VCF)*		1.0	
Sampling Time	2.0 minutes per pump stroke (100 mL)		
Color Change	White → Orange		
Shelf Life	1 year		
Storage condition	Below 10°C (50°F)		
Calibration condition	Calibrated at 20°C (68°F) and 50% RH.		
Reaction Principle	$CHCl_3 + Cr^{6+} + H_2SO_4 \rightarrow Cl_2$		
	CHCl <sub>3</sub> + o - Tolidine → Orange reaction product		
Relative standard deviation	± 10 %		

<sup>\*</sup> Multiply the observed reading by the correction factor (VCF) to obtain the true concentration.

### Possible Interferences:

Compound	Concentration (ppm)	Interference	Colour change / Comments
Cl <sub>2</sub>		+	White to Orange
Br <sub>2</sub>		+	White to Orange
Saturated Halogenated Hydrocarbons		+	White to Orange

### **Correction for Environmental Parameters:**

Temperature	Not necessary between 0 - 40°C (32 - 104°F).	
Relative humidity	Not necessary between 10 - 90 %.	

Manufactured By: Uniphos Envirotronic Pvt. Ltd. Nahuli (PO), Umbergaon (Tal.) Valsad (Dist), Gujarat, India – 396108