# **Uniphos Detector Tube Data Sheet**Direct Reading Length of Stain Type

# Carbon Monoxide co

**SCO-5** 

TWA (TLV): 25 ppm STEL (TLV): 50 ppm Flammable Range: 13-74%

## **Specification:**

Measurement Range	Extended	Standard	Extended	
	50- 1500 ppm	100 - 3000 ppm	200 – 6000 ppm	
No. of Pump Strokes	2(200 mL)	1(100 mL)	1/2 (50 mL)	
Volume Correction Factor (VCF)*	0.5	1.0	2	
Sampling Time	2.5 minutes per pump stroke (100 mL)			
Color Change	White → Brownish green			
Shelf Life	1 year			
Storage condition	Below 10°C (50°F)			
Calibration condition	Calibrated at 20°C (68°F) and 50% RH.			
Reaction Principle	$CO + I_2O_5 + H_2SO_7 \rightarrow I_2$			
Relative standard deviation	± 10 %			

<sup>\*</sup> Multiply the observed reading by the correction factor (VCF) to obtain the true concentration.

### Possible Interferences:

Compound	Concentration (ppm)	Interference	Colour change / Comments
H <sub>2</sub> S	100	No	No Effect
NH <sub>3</sub>	180	No	No Effect
n – Hexane	400	No	Ring
Hydrogen	99.9 %	No	No Effect

### **Correction for Environmental Parameters:**

Temperature	Not necessary between 0 - 40°C (32 - 104°F).	
Relative humidity	Not necessary between 10 - 90 %.	

Manufactured By: Uniphos Envirotronic Pvt. Ltd. Nahuli (PO), Umbergaon (Tal.) Valsad (Dist), Gujarat, India - 396108

Rev.0812