Uniphos Detector Tube Data SheetDirect Reading Length of Stain Type

Carbon Disulphide cs₂

SCS-2

TWA (TLV): 1 ppm Flammable Range: 1.3 - 50%

Specification:

Measurement Range	Extended	Standard	Extended
	1.25 – 25 ppm	2.5 – 50 ppm	5 – 100 ppm
No. of Pump Strokes	4 (400 mL)	2 (200 mL)	0.5 (50 mL)
Volume Correction Factor (VCF)*	0.5	1.0	2.0
Sampling Time	1.5 minutes per pump stroke (100 mL)		
Color Change	Blue → Yellow		
Shelf Life	2 years		
Storage condition	Below 10°C (50°F)		
Calibration condition	Calibrated at 20°C (68°F) and 50 % R.H.		
Reaction Principle	$CS_2 + CrO_3 + H_2S_2O_7 \rightarrow SO_2 + CO_2$		
	$SO_2 + BaCl_2 + H_2O \rightarrow BaSO_3 + 2HCl$		
	HCl + Base → Chloride		
Relative standard deviation	± 10 %		_

^{*} Multiply the observed reading by the correction factor (VCF) to obtain the true concentration.

Possible Interferences:

Compound	Concentration (ppm)	Interference	Colour change / Comments
H ₂ S	50	+	
Ammonia	70	No	No Effect
CH ₄	10000	+	
Hexane	35	No	No Effect

Correction for Environmental Parameters:

Temperature	Not necessary between 0 - 40°C (32 - 104°F).	
Relative humidity	Not necessary between 10 - 90 %.	

Manufactured By: Uniphos Envirotronic Pvt. Ltd. Nahuli (PO), Umbergaon (Tal.) Valsad (Dist), Gujarat, India – 396108