# **Uniphos Detector Tube Data Sheet**

Direct Reading Length of Stain Type

## 1,2-Dichloroethane TWA (TLV): 10 ppm

 $C_2H_4CI_2$ STEL (TLV): NA

Flammable Range: 6.2 - 16.0 %

SDC-2

### **Specification:**

Management	Extended	Standard	Extended
Measurement Range	Do not extend	1 – 50 ppm	Do not extend
No. of Pump Strokes		2 (200 mL)	
Volume Correction Factor (VCF)*		1.0	
Sampling Time	2.5 minutes per pump stroke (100 mL)		
Color Change	White $\rightarrow$ Light Yellow		
Shelf Life	1 year		
Storage condition	Below 10°C (50°F)		
Calibration condition	Calibrated at 20°C (68°F) and 50% RH.		
Reaction Principle	$C_2H_4Cl_2 + CrO_3 + H_2S_2O_7 \rightarrow Cl_2$		
	$Cl_2$ + o-Tolidine $\rightarrow$ Light yellow reaction product		
Relative standard deviation	± 10 %		

\* Multiply the observed reading by the correction factor (VCF) to obtain the true concentration.

#### **Possible Interferences:**

Compound	Concentration (ppm)	Interference	Colour change / Comments
Chlorobenzene	50	+	White to Yellow

### **Correction for Environmental Parameters:**

Temperature	Not necessary between 0 - 40°C (32 - 104°F).	
Relative humidity	Not necessary between 10 - 90 % RH.	

