Uniphos Detector Tube Data SheetDirect Reading Length of Stain Type

1,2-Dichloroethylene C₂H₂CI₂SDH-4M TWA (TLV): 200 ppm STEL (TLV): NA Flammable Range: 5.6 – 12.8 %

TWA (TLV): 200 ppm

Specification:

Measurement Range	Extended	Standard	Extended
	10 – 200 ppm	20 – 400 ppm	40 – 800 ppm
No. of Pump Strokes	2 (200 mL)	1 (100 mL)	0.5 (50 mL)
Volume Correction Factor (VCF)*	0.5	1.0	2.0
Sampling Time	2.0 minutes per pump stroke (100 mL)		
Color Change	White → Orange		
Shelf Life	1 year		
Storage condition	Below 10°C (50°F)		
Calibration condition	Calibrated at 20°C (68°F) and 50% RH.		
Reaction Principle	$C_2H_2CI_2 + CrO_3 + H_2S_2O_7 \rightarrow CI_2$		
	Cl₂ + o - Tolidine → Reaction product		
Relative standard deviation	± 10 %		

^{*} Multiply the observed reading by the correction factor (VCF) to obtain the true concentration.

Possible Interferences:

Compound	Concentration (ppm)	Interference	Colour change / Comments
Trichloroethylene		+	White to Orange
Cl ₂		+	White to Orange
HCI		+	White to Orange

Correction for Environmental Parameters:

Temperature	Not necessary between 0 - 40°C (32 - 104°F).
Relative humidity	Not necessary between 10 - 90 %.

Manufactured By: Uniphos Envirotronic Pvt. Ltd. Nahuli (PO), Umbergaon (Tal.) Valsad (Dist), Gujarat, India – 396108

Rev.0812