Uniphos Detector Tube Data SheetDirect Reading Length of Stain Type

General hydrocarbon (as Hexane) SHE-6L

TWA (TLV): 50 ppm STEL (TLV): NA Flammable Range: 1.1 – 7.5 %

Specification:

Measurement Range	Extended	Standard	Extended
	Do not extend	0.025 - 0.6 %	Do not extend
No. of Pump Strokes		1 (100 mL)	
Volume Correction Factor (VCF)*		1.0	
Sampling Time	1 minutes per pump stroke (100 mL)		
Color Change	Orange → Brownish green		
Shelf Life	2 years		
Storage condition	Below 10°C (50	°F)	
Calibration condition	Calibrated at 20°C (68°F) and 50% RH.		
Reaction Principle	$C_6H_{14} + Cr^{6+} + H_2SO_4 \rightarrow Cr^{3+}$		
Relative standard deviation	± 10 %		

^{*} Multiply the observed reading by the correction factor (VCF) to obtain the true concentration.

Possible Interferences:

Compound	Concentration (%)	Interference	Colour change / Comments
Organic vapours		+	Brownish Green

Other measurable compounds:

Compound	Correction Factor (GCF) *	No. of pump strokes	Measuring range
Hexane	1	1	0.025 - 0.6 %
Cyclohexane	1.3	1	0.0325 - 0.78 %
n-Pentane	0.34	1	0.0085 - 0.204 %
n-Octane	3.8	1	0.095 – 2.28 %
n-Heptane	2.35	1	0.05875 – 1.41 %

^{*} Multiply the observed reading by the correction factor (VCF x GCF) to obtain the true concentration.

Correction for Environmental Parameters:

Temperature	Not necessary between 0 - 40°C (32 - 104°F).
Relative humidity	Not necessary between 10 - 90 %.

Manufactured By: Uniphos Envirotronic Pvt. Ltd. Nahuli (PO), Umbergaon (Tal.) Valsad (Dist), Gujarat, India – 396108