## **Uniphos Detector Tube Data Sheet**Direct Reading Length of Stain Type

# Methylene Chloride TWA (TLV): 50 ppm STEL (TLV): NA

CH<sub>2</sub>CI<sub>2</sub>

SMC-3L

Flammable Range: 15.5 -66.9%

### Specification:

Measurement Range	Extended	Standard	Extended	
	Do not extend	10 – 60 ppm	Do not extend	
No. of Pump Strokes		2 (200 mL)		
Volume Correction Factor (VCF)*		1.0		
Sampling Time	2.5 minutes per pump stroke (100 mL)			
Color Change	White → Pale pink			
Shelf Life	1 year			
Storage condition	Below 10°C (50°F)			
Calibration condition	Calibrated at 20°C (68°F) and 50% RH.			
Reaction Principle	$CH_2CI_2 + Cr^{6+} + H_2SO_4 \rightarrow CI_2$			
	Cl₂ + tetramethylbenzidine → Pale pink reaction product			
Relative standard deviation	± 10 %			

<sup>\*</sup> Multiply the observed reading by the correction factor (VCF) to obtain the true concentration.

#### **POSSIBLE INTERFERENCES:**

Compound	Concentration (ppm)	Interference	Colour change/ Comments
Cl <sub>2</sub>	6	+	White to Pale pink
Trichloroethylene	5	+	White to Pale pink
Methyl Chloride	5	+	White to Pale pink

### **Correction for Environmental Parameters:**

Temperature	Not necessary between 0 - 40°C (32 - 104°F).	
Relative humidity	Not necessary between 10 - 90 %.	



Manufactured By: Uniphos Envirotronic Pvt. Ltd. Nahuli (PO), Umbergaon (Tal.) Valsad (Dist), Gujarat, India - 396108

Rev.0813