# Uniphos Detector Tube Data Sheet Direct Reading Length of Stain Type

# Nitrogen Oxides TWA (TLV): 3 ppm (NO2) & 25 ppm (NO) NOx

SNO-2M

Flammable Range: NA%

## Specification:

Measurement Range	Extended	Standard	Extended
	Do not extend	0.5 - 15 ppm	Do not extend
No. of Pump Strokes		2 (200 mL)	
Volume Correction Factor (VCF)*		1.0	
Sampling Time	1.5 minutes per pump stroke (100 mL)		
Color Change	White $\rightarrow$ Orange		
Shelf Life	1 year		
Storage condition	Below 10°C (50°F)		
Calibration condition	Calibrated at 20°C (68°F) and 50% RH.		
Reaction Principle	$NO + Cr^{6+} + H_2SO_4 \rightarrow NO_2$		
	$NO_2$ + o - Tolidine $\rightarrow$ Yellow reaction product		
Relative standard deviation	± 10 %		

\* Multiply the observed reading by the correction factor (VCF) to obtain the true concentration.

#### **Possible Interferences:**

Compound	Concentration (ppm)	Interference	Colour change / Comments
Cl <sub>2</sub>		+	White to Orange
Br <sub>2</sub>		+	White to Orange
H <sub>2</sub> S	100	No	No Effect
SO <sub>2</sub>	200	No	No Effect

### **Correction for Environmental Parameters:**

Temperature	Not necessary between 0 - 40°C (32 - 104°F).	
Relative humidity	Not necessary between 10 - 90 %.	



Manufactured By: Uniphos Envirotronic Pvt. Ltd. Nahuli (PO), Umbergaon (Tal.) Valsad (Dist), Gujarat, India – 396108

Rev.0812